

## FIT TEST REPORT

11/27/2020

ID NUMBER	1865	CUSTOM1	CN95+
LAST NAME	BABAR	CUSTOM2	
FIRST NAME	BEBI	CUSTOM3	
COMPANY	VENUS	CUSTOM4	
LOCATION	L-17		
TEST DATE	11/27/2020	PORTACOUNT S/N	8038181102
DUE DATE	11/27/2021	N95 COMPANION	Y
RESPIRATOR	VENUS CN95+ CUP [100]	PROTOCOL	OSHA 29CFR1910.134
MANUFACTURER	VENUS	PASS LEVEL	100
MODEL	CN95+		
MASK STYLE	CUP	APPROVAL	NIOSH
MASK SIZE	R	EFFICIENCY<99%	True

<u>EXERCISE</u>	<u>DURATION (sec.)</u>	<u>FIT FACTOR</u>	<u>Pass</u>
NORMAL BREATHING	60	200+	Y
DEEP BREATHING	60	200+	Y
HEAD SIDE TO SIDE	60	200+	Y
HEAD UP AND DOWN	60	200+	Y
TALKING	60	200+	Y
GRIMACE	15	Excl.	
BENDING OVER	60	200+	Y
NORMAL BREATHING	60	200+	Y
OVERALL FF		200+	Y

FIT TEST OPERATOR	_____	DATE	_____
	ASB		
NAME	_____	DATE	_____
	BEBI BABAR		

Note:

### Respirator Fit Test Card

Name: BEBI BABAR	Test Date: 11/27/2020
ID: 1865	Next Test Date: 11/27/2021
<u>Respirator</u>	<u>Results</u>
Mfg: VENUS	Overall FF: 200+
Model: CN95+	FF Pass Level: 100
Style: CUP	Pass: Y
Size: R	Operator: ASB

Protocol: OSHA 29CFR1910.134

Fit Test Method: QNFT using TSI PortaCount

\*\*\* Your company contact information here \*\*\*